



**OPERATING
SECTION
REPORT**



GAS MEASUREMENT MANUAL

**MEASUREMENT OF
GAS PROPERTIES**

PART NO. ELEVEN

A.G.A. GAS MEASUREMENT MANUAL (REVISED)

PART ELEVEN MEASUREMENT OF GAS PROPERTIES

SECTION 11.1—SPECIFIC GRAVITY DETERMINATION
SECTION 11.2—HEATING VALUE OF GAS DETERMINATION
SECTION 11.3—GAS SAMPLING
SECTION 11.4—SULFUR DETERMINATION
SECTION 11.5—TEMPERATURE AND PRESSURE DETERMINATION
SECTION 11.6—SUPERCOMPRESSIBILITY DETERMINATION
SECTION 11.7—DENSITY DETERMINATION
SECTION 11.8—GAS ANALYSIS
SECTION 11.9—GAS ODORIZATION
SECTION 11.10—WATER VAPOR DETERMINATION

**Prepared by the Distribution Measurement
and Transmission Measurement Committee
of the Operating Section**

**American Gas Association
1515 Wilson Boulevard
Arlington, Virginia 22209**

© 1988 AMERICAN GAS ASSOCIATION

All rights reserved

Other Manual Parts Published:

- Part 1 General (XQ1081)
- Part 2 Displacement Measurement (XZ0277)
- Part 4 Gas Turbine Metering (XQ0684)
- Part 5 Other Measurement Methods (XQ0483)
- Part 6 Auxiliary Devices (XQ0779)
- Part 7 Measurement Calculations and Data Gathering (XQ0379)
- Part 9 Design of Meter and Regulation Stations (XQ8803)
- Part 10 Pressure and Volume Control (XQ0584)
- Part 12 Meter Proving (XQ0278)
- Part 13 Distribution Metering Data (XQ0278)
- Part 14 Meter Repair & Selection (XQ0381)

Printed in the U.S.A.

CONTENTS

SECTION 11.1—SPECIFIC GRAVITY DETERMINATION

	Page
Introduction	11.1.1
Definition	11.1.1
Specific Gravity Relationship to Gas Measurements	11.1.1
Effect of Specific Gravity Errors	11.1.1
Gas Sampling	11.1.2
Determining Specific Gravity by Laboratory Methods	11.1.2
Direct Weighing	11.1.2
Gas Analysis	11.1.2
Field Methods for Determining Specific Gravity	11.1.3
Effusion Method	11.1.3
Weighing Method (Gravity Balance)	11.1.3
Centrifugal Method	11.1.7
Kinetic Energy Method	11.1.7
Recording Gravitometers	11.1.9
Recording Weighers	11.1.10
Centrifugal Method	11.1.12
Location and Calibration	11.1.12
Determination of the Specific Gravity of Gas that Contains Distillates	11.1.12
References	11.1.13

SECTION 11.2—HEATING VALUE OF GAS DETERMINATION

	Page
Terms and Definitions	11.2.1
Calculating the Heating Value from Gas Analysis	11.2.2
Total Heating Value Instruments	11.2.3
Recording Calorimeter	11.2.3
Thermeter	11.2.6
Stoichiometric Combustion Calorimeter	11.2.7
Special Precautions	11.2.10
Chromatographs	11.2.10
Sample System	11.2.11
Process Chromatograph	11.2.12
Controller	11.2.13
Correction of Gas Heating Value for Water Vapor Content	11.2.13
Introduction	11.2.13
Water Vapor in Saturated Gas	11.2.14
Adjusting BTU for Water Vapor	11.2.16
Pressure Base	11.2.16
Temperature Base	11.2.16
Water Vapor Base	11.2.16
Saturated to Dry Conversion	11.2.16
Dry to Saturated Conversion	11.2.17
Dry to Actual Water Vapor Conversion	11.2.17
Saturated to Actual Water Vapor Content Conversion	11.2.17

SECTION 11.3—GAS SAMPLING

	Page
Introduction	11.3.1
Sampling Probes	11.3.1
Continuous On-Line Sampling	11.3.2
Spot Sampling	11.3.4
Accumulative Sampling Over a Period of Time	11.3.4
Cleaning and Filling the Sample Cylinder	11.3.5
Purging Schedule for the Sample Cylinder	11.3.5
Discharging of Sample Containers: The Complete Fluid Vaporization Method	11.3.6
Sampling Materials	11.3.8
Determination of Water Vapor and Hydrogen Sulfide	11.3.8
Shipping of Samples	11.3.8

SECTION 11.4—SULFUR DETERMINATION

	Page
Introduction	11.4.1
Lead Acetate Paper	11.4.2
Colorimetric Gas Detectors	11.4.2
The Cadmium Sulfate Test	11.4.3
Methylene Blue Method	11.4.3
Total Sulfur Determination by Hydrogenation	11.4.4
Electrolytic Titrators	11.4.5
H ₂ S Analyzers	11.4.8
Chromatographs	11.4.8
References	11.4.12

SECTION 11.5—TEMPERATURE AND PRESSURE DETERMINATION

TEMPERATURE DETERMINATION

	Page
Definitions	11.5.1
Temperature	11.5.1
Absolute Temperature	11.5.2
Temperature Recorders	11.5.2
Filled-System Temperature Recorders	11.5.2
Liquid-Filled Thermal System (Class I)	11.5.3
Vapor Pressure Thermal System (Class II)	11.5.4
Gas-Filled Thermal System (Class III)	11.5.5
Mercury-Filled Thermal System (Class IV)	11.5.5
General	11.5.5
Installation	11.5.7
Operation and Testing	11.5.7
The Chart	11.5.7
Test and Inspection of the Recorder	11.5.7
Field Test Procedure	11.5.8
Spot Checks	11.5.8
Constant Temperature Condition	11.5.8
Varying Temperature Conditions	11.5.8

PRESSURE DETERMINATION

Definitions	11.5.9
Atmospheric Pressure	11.5.9
Absolute Pressure	11.5.9
Gauge Pressure	11.5.10
Pressure Measuring Devices	11.5.10
Mechanical Pressure Gauges	11.5.10
Bourdon Tube	11.5.10
Diaphragm and Bellows	11.5.10
Tests and Inspection of Pressure Gauges	11.5.10
Deadweight Instruments	11.5.10
Deadweight Gauge	11.5.10
Deadweight Tester	11.5.12
Some Basic Rules for Use of Deadweight Instruments	11.5.12
Manometers	11.5.13
General	11.5.13

SECTION 11.6—SUPERCOMPRESSIBILITY DETERMINATION

	Page
The Ideal Gas Law	11.6.1
Deviation from the Ideal Gas Law	11.6.1
Supercompressibility Determination by Test	11.6.3
Supercompressibility Manuals	11.6.4
References	11.6.6

SECTION 11.7—DENSITY DETERMINATION

	Page
Introduction	11.7.1
Definitions	11.7.1
Density	11.7.1
Specific Weight	11.7.1
Mass(m)-Weight(w)	11.7.1
Engineering Commentary	11.7.1
Gas Density Measurement Methods	11.7.2
Direct Methods	11.7.2
Indirect Methods	11.7.2
Criteria for Selecting Densitometers	11.7.2
Installation and Practice	11.7.2
Examples	11.7.5
Turbine Flowmeter Run (Typical)	11.7.6
Orifice Meter Run (Typical)	11.7.6
Primary Uses of Densitometers	11.7.6
Field Calibration of Densitometers	11.7.7
The Pycnometer	11.7.7
Secondary Transfer Standard for Calibrating On-Line Densitometers	11.7.9
Gas Laws Applied to Calibration	11.7.10
Mathematics of Gas Mass Density Measurement and Flow Metering	11.7.10
Specific Gravity (G)	11.7.12

Orifice Flow Equation	11.7.12
Base Volume Flow from Weight Flow and G	11.7.13
Flow Equations for Turbine, Rotary, and Positive Displacement Meters	11.7.13
Flow System Coefficient Optimization-Utilizing the Mass Density Function	11.7.14
Other Applications Related to Orifice Plate Flow Coefficient Correction	11.7.14
Turbine Flowmeter K Factor Corrections	11.7.15
Therms (Energy Metering)	11.7.16
Therms Flow/Mass Flow	11.7.16
Conclusions	11.7.16
References	11.7.16

SECTION 11.8—GAS ANALYSIS

	Page
Introduction	11.8.1
Chromatography	11.8.1
Water Vapor and Sulfur Determination	11.8.1
Volumetric Analysis	11.8.1
Theory	11.8.1
Apparatus	11.8.2
Accuracy	11.8.3
Analysis of CO ₂ by Infrared Absorption	11.8.3
Micro-Fuel Cell Analysis of Trace Oxygen	11.8.3

SECTION 11.9—GAS ODORIZATION

	Page
Introduction	11.9.1
Required Odor Intensity	11.9.1
Explosive Limits	11.9.1
Selection of an Odorant	11.9.1
Required Properties	11.9.1
Desirable Properties of an Odorant	11.9.2
Commonly Used Odorant Compounds	11.9.2
Mercaptan Odorants	11.9.2
Odorizing Equipment	11.9.4
Types of Odorizers	11.9.4
Liquid Injection Odorizers	11.9.4
Evaporation Odorizers	11.9.7
Odorizer Installation	11.9.9
Site Selection	11.9.9
Materials of Construction	11.9.9
Vapor Control	11.9.9
Handling Odorant in Drums	11.9.10
Returnable Cylinders	11.9.10
Handling of Spills	11.9.10
Odor Level Determination	11.9.11
Dilution Apparatus	11.9.11
Room Test	11.9.11
Sniff Test	11.9.13

Instrumentation Methods of Odorant Analysis	11.9.14
Electrolytic Titration	11.9.14
Gas Chromatography	11.9.15
Colorimetric Methods	11.9.15
Detector Tubes	11.9.15
Wet Chemical Technique	11.9.16

SECTION 11.10—WATER VAPOR DETERMINATION

	Page
Introduction	11.10.1
General	11.10.1
Sampling	11.10.1
Concentration	11.10.1
Location of Sampling Device	11.10.6
Selection of Materials for Sampling Systems	11.10.6
Volume of the System	11.10.6
Constant Pressure	11.10.6
Sample Site Selection	11.10.6
Methods of Determining Water Vapor Content	11.10.6
Dew Point or Condensate Method	11.10.6
Electrolytic Water Analyzer	11.10.8
Direct Chemical Method	11.10.9
Hygroscopic Conductivity	11.10.10
Quartz Crystal Oscillator	11.10.11
Aluminum Oxide Hygrometer	11.10.11
Calibration of Water Vapor Measuring Instruments	11.10.11
Dew Point Method	11.10.11
Chemical Methods	11.10.12
Permeation Tubes	11.10.12
Other Methods	11.10.12
Bibliography	11.10.12

FIGURES

Figure		Page
11.1.1	Portable specific gravity balance	11.1.5
11.1.2	Gas gravity balance pressure ratio adjustment values to compensate for Doyle's law deviation	11.1.5
11.1.3	Factors for converting inches of mercury at manometer temperature to inches of mercury at 32°F	11.1.6
11.1.4	Schematic of gas centrifuge gravitometer	11.1.7
11.1.5	Cross-section of the centrifuge wheel	11.1.8
11.1.6	Operating principle—portable indicator	11.1.9
11.1.7	Schematic diagram of a recording gravitometer having gas-filled drums suspended from a balance beam	11.1.10
11.1.8	A recording gravitometer in which the unit weight of a column of gas is compared to a unit weight of an equal column of air	11.1.11
11.1.9	Schematic diagram of a recording gas gravitometer with single float on a weight-opposed beam	11.1.12
11.2.1	Flow diagram and part identification of a recording calorimeter	11.2.5
11.2.2	Schematic flow diagram of a thermeter	11.2.7
11.2.3	Functional diagram of a stoichiometric combustion calorimeter	11.2.9
11.2.4	Functional overview of a gas BTU transmitter	11.2.9
11.2.5	Gas analysis system using chromatograph	11.2.11
11.2.6	Process chromatograph	11.2.12
11.2.7	Water content of actual gas versus temperature	11.2.15
11.3.1	Sampling stinger	11.3.2
11.3.2	Piping connection for calorimeter	11.3.4
11.3.3	Purging schedule and equations for the sample cylinder	11.3.6
11.3.4	Typical constant pressure sampler	11.3.7
11.4.1	Physical effects of inhaling H ₂ S	11.4.1
11.4.2	Two types of gas detectors	11.4.3
11.4.3	Schematic of cadmium sulfate test	11.4.4
11.4.4	Absorption flask	11.4.4
11.4.5	Schematic diagram of apparatus for sulfur method	11.4.5
11.4.6	Schematic of an electrolytic titrator	11.4.6
11.4.7	Schematic of an electrolytic titration cell	11.4.6
11.4.8	Typical chart of an electrolytic titrator	11.4.8
11.4.9	Schematic of an H ₂ S analyzer	11.4.9
11.4.10	Typical chart of a continuous recording H ₂ S analyzer	11.4.10
11.4.11	Typical chart of a fixed-cycle recording H ₂ S analyzer	11.4.10
11.4.12	Schematic of a gas chromatograph	11.4.11
11.5.1	Temperature scales	11.5.1
11.5.2	The filled thermal system of a typical temperature recorder	11.5.3
11.5.3	A temperature recorder system fully compensated for the effects of ambient temperature	11.5.4
11.5.4	A case compensated thermal system	11.5.4
11.5.5	A Class II-A thermal system	11.5.6
11.5.6	A Class II-B thermal system	11.5.6
11.5.7	A Class II-C thermal system	11.5.6
11.5.8	A Class II-D thermal system	11.5.6
11.5.9	Diagram of relations between absolute and gauge pressures	11.5.9
11.5.10	Bourdon tube and helical	11.5.11
11.5.11	Diaphragm and bellows elements	11.5.11
11.5.12	Deadweight gauge	11.5.11
11.5.13	Deadweight tester	11.5.13

11.6.1	Variation of compressibility factor with pressure for three general types of gas	11.6.3
11.6.2	Pure hydrocarbon gas 0.600 specific gravity	11.6.7
11.7.1	Insertion type densitometer—direct into flowing pipeline	11.7.3
11.7.2	Sampling type densitometer—the sample lines straddle a nonflow measuring orifice inducing flow through the densitometer	11.7.3
11.7.3	Sampling type densitometer—sampling lines are connected to positive and negative pressure sample tubes to induce flow through the densitometer	11.7.4
11.7.4	Sampling type densitometer—employs sampling blower to induce flow through densitometer at rates independent of flow rate in main line stream	11.7.4
11.7.5	Point of density measurement to flowmeter element—for orifices, point 2 is the static pressure tap; for turbine meters point 2 is upstream of the rotor	11.7.5
11.7.6	Typical turbine meter run showing alternate positions relative to turbine meter run for insert-type densitometer	11.7.6
11.7.7	Typical turbine meter showing sampling type densitometer relative to turbine meter run	11.7.7
11.7.8	Typical orifice meter run showing four alternate positions for densitometers	11.7.7
11.7.9	Field calibration, with pycnometer or master densitometer, for typical line sampling densitometer	11.7.8
11.7.10	Field calibration system, with pycnometer or master densitometer, for typical in-line densitometer	11.7.9
11.7.11	Field calibration system for typical in-line densitometer using an in-line master densitometer	11.7.10
11.7.12	Field calibration of typical in-line densitometer by computation from gas law equation	11.7.11
11.8.1	Orsat volumetric analysis apparatus	11.8.2
11.9.1	Drip odorizer for liquid injection	11.9.5
11.9.2	Typical diaphragm displacement pump with pneumatic controller	11.9.6
11.9.3	Single-unit wick odorizer	11.9.8
11.9.4	By-pass odorizer	11.9.8
11.9.5	Three views of a vehicle used for conducting the gas odorant room test	11.9.12
11.10.1	Water vapor content of compressed natural gas	11.10.3
11.10.2	Bureau of Mines dewpoint tester	11.10.5
11.10.3	Thermoelectric dewpoint apparatus	11.10.7
11.10.4	Schematic of electrolytic water analyzer	11.10.8
11.10.5	Automatic direct chemical titration method	11.10.10

TABLES

Table	Page
11.1.1 Sample Specific Gravity Determination	11.1.3
11.2.1 Physical Properties of Gases at Exactly 14.73 Pounds per Square Inch Absolute and 60°F	11.2.3
11.2.2 Sample Calculation of Heating Value from Gas Analysis	11.2.4
11.2.3 Summation Factors at 60°F (288.71K)	11.2.4
11.3.1 Natural Gas (Typical Casing-Head) with High Boil-Point Hydrocarbons	11.3.8
11.4.1 Component Analysis	11.4.8
11.4.2 Range of Measurement by Type of Detector	11.4.11
11.10.1 Typical Vapor Pressure of Water	11.10.2

Section 11.1

SPECIFIC GRAVITY DETERMINATION

INTRODUCTION

Definition

Specific gravity is a term used to express the ratio of the molecular weight of a gas (or a mixture of gases) to the molecular weight of air. Air is a compound of varying composition. Its molecular weight has been established as 28.9625, as reported in the Table of Physical Properties, A.G.A. Report #3, revised 5/16/85. Where precise specific gravity determinations are made by analytical means, the prevailing composition of air is of no consequence if consistently related to the number 28.9625. Yet, considering technical inability to determine the molecular weight of gas short of analytical fractionation, the ratio of gas density to the density of air under common pressure and temperature conditions is often taken to represent specific gravity. The difference between relative density and relative molecular weight is small, provided, however, that densities are determined at relatively low pressures. Any difference between the two ratios may be attributed to the difference in Ideal Gas Laws deviation at any given common pressure and temperature, if we can assume the composition of reference air at a particular locale approximates a molecular weight of 28.9625. Relative densities may be converted to relative molecular weight (specific gravity) by accounting for the Ideal Gas Laws.

Specific Gravity Relationship to Gas Measurements

Specific gravity is a fundamental aspect of a gas in several gas measuring sectors. In flow metering, the pressure drop across an orifice, flow nozzle, venturi tube, or similar restriction is affected by the molecular weight of the flowing fluid. Also, the relationships that have been found between specific gravity of a gas and its deviation from Ideal Gas Laws account for the preparation of A.G.A.'s *Manual for the Determination of Supercompressibility Factors for Natural Gas* PAR Research Project NX-19*. Similar relationships have also been found between specific gravity and the calorific value of fuel gases and are covered by A.G.A.'s publication Report No. 5, *Fuel Gas Energy Metering*.

Effect of Specific Gravity Errors

Any error in specific gravity determination is very significant, if examined in the light of any one of the several factors affected— F_{gr} , F_{pv} , F_{hgm} , and the heating value of a gas. Yet the overall effect of specific gravity error will be found less than the sum of individual errors, considering the error attributed to a single factor is often offset by compensating errors in other factors tied to a specific gravity index. The effect of specific gravity error on individual factors and combined factors used in orifice meter measurements is illustrated in the following example for a hydrocarbon gas mixture flowing at 1000 psia, at 60°F, under standard gravitational forces of 980.665 cm/sec./sec. (32.174 ft/sec²) at sea level and 45° latitude.

EXAMPLE

	Specific Gravity	Factor F_{gr} (Report #3)	Factor F_{pv} (NX-19)	Factor F_{hgm} (Report #3)	Btu/cf (Report #5)	Combined Factors
Correct	.600	1.2910	1.0834	.9978	1086.9	1516.9
Erroneous	.610	1.2804	1.0872	.9977	1102.6	1531.3
% Error	+1.67	−0.82	+0.35	−0.01	+1.44	+0.95

*This manual has been superseded by A.G.A. Transmission Measurement Committee Report #8. A.G.A. 8 extends the range of contaminants, pressures, and temperatures allowed when it is used to calculate F_{pv} .

If the gas used in the example were other than a pure hydrocarbon mixture, contained diluents would have changed F_{pv} factors, the heating value of the gas, and the combined factors shown. The direction of change depends on the concentration of individual diluents in the gas mixture.

Specific gravity error has a greater effect on displacement metering than will be found with orifice metering, considering that factors F_{gr} and F_{hgm} do not apply to displacement meters and that the effect of the supercompressibility factor is roughly twice as large $-(F_{pv})^2$. All errors arising from specific gravity determinations are additive.

Gas Sampling

Aside from the technique employed in specific gravity determinations, gas sampling practices must be emphasized in the interest of accuracy. Gases diverted to sample bottles or gravimeters should be withdrawn from the center of a flowing stream through a probe whose inlet is near midpoint of the runway cross section. In the absence of a probe or stinger, condensibles in a gas stream will creep along pipewalls to the outlet port and thence into bottles or instruments, in which case a high specific gravity reading may be expected.

Other precautions must be taken where gases contain hydrogen sulfide, carbon dioxide and other constituents which will react with, or be absorbed by, such materials as iron, steel, brass, or rubber. Sampling conduits of aluminum or stainless steel will be found suitable for most gas sampling applications.

DETERMINING SPECIFIC GRAVITY BY LABORATORY METHODS

Direct Weighing

The direct weighing method for finding the specific gravity of a gas represents one of two methods adjudged most accurate although seldom used in practice. Here, an analytical balance is used. Two glass bulbs of matched volume and weight are filled, one with gas and the other with dry air. Both samples must be maintained at equal temperature and pressure. The bulbs are then placed on the balance and weighed. The procedure is repeated, except that the samples are exchanged in the bulbs after thorough purging. Again the weights are observed and recorded, and the average of the two observations is used in determining the weight of each. The ratio of the average weight of the gas sample to the average weight of the air sample is the specific gravity. The procedure of changing sample bulbs has a compensating effect on difference in bulb weight, bulb buoyancy, and slight changes in pressure and temperature. This method is used in laboratories under ideal conditions by skilled technicians using precision instruments; consequently, it does not lend itself to field use.

Gas Analysis

Specific gravity may be determined from a gas analysis obtained by fractional distillation or a mass spectrometer but preferably with a gas chromatograph. Properly conducted, the analysis will provide a quantitative breakdown of the composition from which the specific gravity may be calculated. The specific gravities of the individual components of a gaseous mixture are taken from a reference table, such as is found in A.G.A. Measurement Committee Report No. 3, and multiplied by the mole fraction of each component. The sum of the products is the ideal specific gravity of the gas on a dry basis. If the gas is saturated with water vapor, an adjustment should be made for greater accuracy; however, since this correction is quite small, it is ordinarily disregarded.

Pipeline gas may contain constituents in vapor phase that the chromatograph is not programmed to analyze, such as methanol, amines, glycols, compressor lubricating oils, scrubber oil, and various types of inhibitors. The quality and vapor pressure of these unanalyzed components will affect the calculated specific gravity if the components are in appreciable concentration.

Table 11.1.1 gives an example of specific gravity determination from a fractional gas analysis:

TABLE 11.1.1
Sample Specific Gravity Determination

Compound	A Compound Mole Fraction (An Example)	B Specific Gravity, Pure Compounds (Constants)*	A × B
Hydrogen	Nil	.06960	—
Helium	.0001	.13820	.00001
Water Vapor	Nil	.62202	—
Carbon Monoxide	Nil	.96711	—
Nitrogen	.0040	.96723	.00387
Oxygen	Nil	1.10484	—
Hydrogen Sulfide	Nil	1.17669	—
Argon	Nil	1.37930	—
Carbon Dioxide	.0050	1.51955	.00760
Methane	.9312	.55392	.51581
Ethane	.0200	1.03824	.02076
Propane	.0101	1.52256	.01538
Iso-Butane	.0088	2.00684	.01766
N-Butane	.0070	2.00684	.01405
Iso-Pentane	.0069	2.49115	.01719
N-Pentane	.0053	2.49115	.01320
Hexanes (avg.)	.0010	2.97547	.00298
Heptanes (avg.)	.0004	3.45978	.00138
Octanes (avg.)	.0002	3.94410	.00079
Mixture	1.0000		0.63068

*From A.G.A. Report No. 3, Revised 5/16/85.
All values are ideal, at 14.73 psia, 60°F.

(Ideal Specific
Gravity, uncor-
rected for "Z")

FIELD METHODS FOR DETERMINING SPECIFIC GRAVITY

Effusion Method

The effusion method is based on the principle that the specific gravity of a gas is equal to the ratio of the square of the time required for a definite volume of gas to pass through a given orifice as compared to the square of time required for an equal volume of air to pass through the same orifice under the same conditions of temperature and pressure. While the theory of this method is sound technically, it is usually associated with Shilling Apparatus, where water is used as a gas drive to expel the gases to atmosphere through a small orifice. Engrained water vapor, condensation about the orifice, and the necessity to use a stop watch to measure time have discredited the method.

Weighing Method (Gravity Balance)

Indirect weighing is used extensively in the natural gas industry, primarily because of the portability of the equipment. The instruments used employ two principles: **Archimedes Principle**, which states that the buoyant force exerted on a body suspended in a fluid is proportional to the density of the fluid, and **Ideal Gas Laws**, which state that the density of a gas is proportional to its absolute pressure at a given temperature. Two types of indirect weighing or gravity balances are used. In the first type (Figure 11.1.1), a balanced

condition of the float and counterweight beam is obtained by adjustment of the pressure of the gas or air in the chamber. Equal gas and air densities are used to balance the beam with the assumption that the ratio of gas to air pressure is indicative of the relative density of the two substances following Boyle's Law. With the second type, gas and air chamber pressures are common (usually atmospheric). Relative density is indicated by the difference in the angles of deflection of the beam, comparing the weight of gas to air. Of the two indirect weighing methods, the first type is used most commonly. In either instance, relative densities are found. These relative densities may be converted to ideal specific gravity (molecular weight ratio) by accounting for the ratio of Ideal Gas Laws deviation of gas to that of air. In the absence of precise data relating to Boyle's Law deviation of the gas and air at measured pressures and temperatures, Figure 11.1.2 may be used in conjunction with the first named method.

$$\text{Specific Gravity} = \frac{(P_g \times T_g \times Z_g)}{(P_a \times T_a \times Z_a)}$$

where

P=absolute pressure

T=thermodynamic temperature (Kelvin or Rankine)

Z=Compressibility factor

Subscript g= gas,

Subscript a= air.

In the instance of the second indirect weighing method (where the gas and the air are at equal pressure and temperature and near one atmosphere and 60°F temperature), the Ideal Gas Laws deviation ratio may be expressed as $1.00369 - 0.0101 \text{ RD}$, where RD is relative density. Thus

$$\text{Specific Gravity} = \text{RD} (1.00369 - 0.0101 \text{ RD})$$

Auxiliary requirements for making a determination of specific gravity with a gravity balance are a pressure-suction pump, an air drier, a mercury manometer, and a mercurial or aneroid barometer. For field use, a temperature-compensated aneroid barometer is used normally. It should be adjusted to read in inches (or millimeters) of mercury at 32°F. However, it must not be adjusted for elevation or latitude since it must indicate the actual atmospheric pressure at the place of use. An aneroid barometer should be checked frequently against a reference standard mercurial barometer. Indicating silica gel is a practical drying agent for the air drier because it changes color from blue to pink coincident with the absorption of moisture.

When one conducts field determinations with a specific gravity balance, the following precautions should be observed. The instrument should be located where it will be free from vibration and wind currents, and it should be on a firm base and remain level throughout the test. It should be in the test location long enough to attain equilibrium with the ambient temperature before a test is starting; the ambient temperature and the sample temperature should be nearly equal. Sample connections should be purged thoroughly and care should be taken to ensure that no liquid droplets enter the apparatus. To avoid condensation of vapors, the temperature of the balance should be higher than that of the supply gas. Care should be taken to obtain air samples from an uncontaminated source.

The procedure for determining specific gravity with this instrument is as follows. Preferably, the balance is mounted on a tripod or other firm stand then leveled after being connected to the gas supply, the manometer, air drier, and pump. All connections should be checked for leaks against both positive and negative gauge pressures. Next, observe and record the barometer reading. With the suction side of the pump connected to the balance, exhaust the air, then admit air slowly through the drier. Twice again exhaust and refill through the drier. The last refilling operation should be accomplished with the balance beam unlocked, slowly admitting air through the drier until the balance beam reaches the balance or equilibrium position. It must be observed that the beam is in a free equilibrium position, indicated by equal oscillations. Next, lock the beam and read the manometer after tapping it lightly; then, note and record




Figure 11.1.1 Portable specific gravity balance.




Figure 11.1.2 Gas gravity balance pressure ratio adjustment values to compensate for Boyle's law deviation.

5800019

the temperature in the balance chamber. Subsequently, the air should be exhausted from the balance and then gas admitted slowly until the pressure in the case is about 25" of mercury above atmospheric pressure. The gas to be tested is not passed through the drier. Repeat the exhausting and repressuring operations with gas at least two or three times. Unlock the balance beam and release gas slowly over the last refill operation until the beam again reaches the equilibrium position—be certain that it is free, as in the case of the air. Again, lock the beam, tap and read the manometer, and read the temperature within the balance. It is advisable to check the air again, considering that there are minor changes in atmospheric air from time to time. The second air pressure reading should agree closely with the first (2.0 mm of Hg difference is acceptable). Record the temperatures of the balance and the barometer. Then average the two barometer and air-pressure readings, and then compute the absolute pressures of the air and gas. Be sure that the readings of both the manometer and barometer are taken at equal temperature. If an aneroid barometer is used, readings must be adjusted to provide inches of mercury at 32°F by application of a correction factor which may be read from Figure 11.1.3. If a mercury barometer is used and the temperatures of the barometer and mercury manometer are essentially the same, no temperature corrections need be applied to either. However, both are subject to adjustment if they differ from each other in temperature.

Ordinarily, the absolute pressure of air in the balance will be less than atmospheric, while that of gas will be greater. If the mean of the two air temperatures differs from the gas temperature by 2°F or more, the specific gravity should be corrected for such difference by multiplying it by the inverse ratio of the absolute temperatures. If this temperature difference exceeds as much as 5°F, test results should be discarded and another test should be conducted. If the specific gravities calculated from two tests do not agree within 0.002–0.003, additional tests should be made until two are found to agree within tolerance.

The second type of gravity balance is somewhat similar to the first, except that its is designed for use at atmospheric pressure only, with the inclinations of the balance beam being observed under atmospheric pressure within the weighing chamber.




Figure 11.1.3 Factors for converting inches of mercury at manometer temperature to inches of mercury at 32°F.

Centrifugal Method

Figure 11.1.4 is a schematic presentation of an indicating gravitometer that employs centrifuge principles. Its mode of operation is similar to that of a centrifugal water pump, having the capability of creating a high gas head pressure over a short operating radius. Here a single wheel copes with both reference air and the gas under observation (see Figure 11.1.5). Good thermal conductivity insures temperature equality. Pressures within the rotor are measured through lances inserted in inlet orifices. There is no requirement for barometric pressure compensation, since the gravitometer is inherently self-compensating. Two manometers are used to detect head pressures. In practice, the speed of wheel rotation is adjusted to cause the air pressure to read 1.0 on the air manometer, which represents the specific gravity of air. The specific gravity of the observed gas is read directly from a scale that indicates the column height of the fluid in the sample gas manometer.

Kinetic Energy Method

Instruments used to determine specific gravity by the Kinetic Energy method use the kinetic energy equation $KE = mv^2 \div 2$. They determine the specific gravity by means of the difference in torque produced by




Figure 11.1.4 Schematic of gas centrifuge gravitometer.