



G. M. S. de Silva

BASIC METROLOGY *for* ISO 9000 CERTIFICATION

Basic Metrology for ISO 9000 Certification

G.M.S. de Silva

 **Routledge**
Taylor & Francis Group
LONDON AND NEW YORK

First published by Butterworth-Heinemann

This edition published 2011 by Routledge
2 Park Square, Milton Park, Abingdon, Oxfordshire OX14 4RN
711 Third Avenue, New York, NY 10017, USA

First issued in hardback 2016

Routledge is an imprint of the Taylor and Francis Group, an informa business

First published 2002

© G.M.S. de Silva 2002

All rights reserved. No part of this publication may be reproduced in any material form (including photocopying or storing in any medium by electronic means and whether or not transiently or incidentally to some other use of this publication) without the written permission of the copyright holder except in accordance with the provisions of the Copyright, Designs and Patents Act 1988 or under the terms of a licence issued by the Copyright Licensing Agency Ltd, 90 Tottenham Court Road, London, England W1P 9HE. Applications for the copyright holder's written permission to reproduce any part of this publication should be addressed to the publishers

British Library Cataloguing in Publication Data

A catalogue record for this book is available from the British Library

Library of Congress Cataloguing in Publication Data

A catalogue record for this book is available from the Library of Congress

Typeset at Replika Press Pvt Ltd, Delhi 110 040, India

ISBN 13: 978-1-138-17222-7 (hbk)

ISBN 13: 978-0-7506-5165-3 (pbk)

Basic Metrology for
ISO 9000 Certification

Foreword

Metrology is often considered to be simply a field of science and technology concentrating on the measurement and accuracy of things we make. However, much of the historical development of technology suggests this chicken and egg situation is reversed, in that evolving the ability to measure qualities more accurately allows us to develop more reliable ways of manufacturing things, as we now have a means of controlling quality.

The recent rise in nano-technological advances is a beacon to this premise, which was recognized in an earlier era of technological development, that might be called the micro-technology phase associated with the industrial revolution. In particular by prominent people of that era, such as one of the founders of modern metrology, Joseph Whitworth, who in his Presidential Address to The Institution of Mechanical Engineers, in London in 1856, said the following;

“I would next call your attention to the vast importance of attending to the two great elements of constructive mechanics, namely, the true plane and the power of measurement”,

and

“I hope the members of this institution will join me in doing what we can with reference to these two important subjects – correct measurement and its corollary proper graduation of size. The want for more correct measurement seems to pervade everything.”

Little has changed in our needs in this respect, and the importance of metrology to the advancement of science and technology remains, albeit on improved levels of accuracy.

The author, Swinton de Silva, has worked for the majority of his working life, a period covering several decades, in this field, and the impressive breadth of his metrology knowledge is reflected in the very broad ranging contents of this book. As well as representing an extremely valuable contribution to the literature in the metrology field, this book also represents an essential guide to those involved with ISO 9000 Certification.

Dr R S Sayles
Reader in Mechanical Engineering
Imperial College of Science, Technology and Medicine
London, UK

Preface

Test and measuring instruments are used extensively in the modern manufacturing and processing organizations. The accuracy of the measurements made by these instruments has a direct impact on the quality of the product or service provided by the organization. The recent developments in the field of certified quality and environmental management systems, namely registration to ISO 9000 and ISO 14000 standards, require that test and measurement equipment be periodically calibrated using measurement standards traceable to the international measurement system. In addition there is also the necessity for test and calibration laboratories to be accredited by a third party certification body in accordance with the international standard ISO/IEC 17025 (previously ISO/IEC Guide 25).

Although a number of books are available describing specific measurement fields such as temperature and pressure, books covering a number of important measurement fields are few. This book intends to fill this gap. The book is primarily aimed at persons working in industry whose duties are related to calibration and maintenance of test and measurement equipment. Students reading for bachelor's degrees or diplomas in the fields of electrical, mechanical and production engineering and related technology-based courses can also use it as an introduction to metrology.

The book is an introduction to fundamental measurement principles and practical techniques used in the calibration of test and measuring equipment belonging to seven measurement fields, namely length, angle, mass, temperature, pressure, force and electrical metrology. Fundamental concepts of measurement and calculation of measurement uncertainties are also dealt with.

G M S de Silva

Acknowledgements

The author wishes to acknowledge gratefully the many persons who assisted him during the course of the preparation of the book. In particular Professor Parry Jones of the King Fahd University of Petroleum and Minerals, Dhahran, Saudi Arabia, Dr R.S. Sayles of the Imperial College of Science, Medicine and Technology, University of London, D.R. White and M. Clarkson of the Measurement Standards Laboratory of New Zealand, and A. Maluyo and N. Liyanage for reviewing several chapters of the book and making valuable suggestions.

The assistance provided by Chandimal Fernando by drawing a large number of illustrations, Ms Roshini Thoradeniya by editing the script, Sunil Amarawansa, Kishantha Galappatti, Tharana Thoradeniya and Ms Imitha de Silva are gratefully acknowledged.

I am also grateful to the Institute of Measurement and Control, UK, the National Physical Laboratory, UK, the National Institute of Advanced Industrial Science and Technology, Japan, the National Metrology Institute of Japan, the National Institute of Standards and Technology (NIST), USA, the Commonwealth Scientific and Industrial Research Organization of Australia, the National Research Council of Canada, The International Organization for Standardization, the International Electrotechnical Commission, the International Organization for Legal Metrology, the American National Standards Institute, Fluke Corp., USA, Mitutoyo Corp., Japan, Isothermal Technology Ltd. UK, Morehouse Instrument Co. USA and Hottinger Baldwin Measurements (HBM) of Germany for providing materials and granting permission to reproduce them in the book.

Figures 3.3, 3.4, 3.5 and Table 3.1 (Critical characteristics of gauge blocks taken from ISO 3650: 1998), Table 6.5 (Classification of force proving instruments from ISO 376: 1987) and Table 6.7 (Classes of testing machines from ISO 7500-1: 1999) have been reproduced with the permission of the International Organization for Standardization, ISO. These standards can be obtained from any ISO member body or directly from the Central Secretariat, ISO, Case Postal 56, 1211 Geneva 20, Switzerland. Copyright remains with ISO.

The author thanks the International Electrotechnical Commission (IEC) for permission to reproduce extracts from IEC publications IEC 60751 (1983-01) entitled 'Industrial platinum resistance thermometer sensors' and IEC 60584-2 (1982-01) entitled 'Thermocouples. Part 2: Tolerances'. All extracts are the copyright of IEC, Geneva, Switzerland. All rights reserved. Further

information on the IEC, its publications and its activities is available from www.iec.ch. IEC has no responsibility for the placement and context in which the extracts and contents are reproduced in this publication; nor is IEC in any way responsible for any of the other content or accuracy of this publication.

Contents

<i>Foreword</i>	<i>ix</i>
<i>Preface</i>	<i>xi</i>
<i>Acknowledgements</i>	<i>xiii</i>
1 Requirements of ISO 9000 standards for test and measuring equipment	1
1.1 Introduction	1
1.2 Evolution of ISO 9000 standards	1
1.3 Requirements of ISO 9001: 2000	2
Bibliography	5
2 Fundamental concepts of measurement	7
2.1 Introduction	7
2.2 Fundamental concepts	8
Bibliography	16
3 Linear and angular measurements	17
3.1 Introduction	17
3.2 Length measurement	17
3.3 Calibration of dimensional standards and measuring instruments	31
3.4 Measurement of angle	40
Bibliography	44
4 Mass measurements	45
4.1 Introduction	45
4.2 Primary standard and SI units	45
4.3 Secondary and working standards	45
4.4 Mass and weight	45
4.5 Mass standards – types and classes	49
4.6 Industrial weighing systems	60
4.7 Accuracy classes of balances	61
4.8 Calibration of balances	62
Bibliography	67

5	Pressure measurements	69
5.1	Introduction	69
5.2	SI and other units	69
5.3	Absolute, gauge and differential pressure modes	70
5.4	Primary standards	70
5.5	Spinning ball gauge standard	73
5.6	Secondary standards	74
5.7	Working standards	74
5.8	Pressure measuring instruments	81
5.9	Calibration of pressure standards and instruments	91
	Bibliography	99
6	Measurement of force	101
6.1	Introduction	101
6.2	SI and other units	101
6.3	Primary standard	102
6.4	Secondary standards	102
6.5	Force measuring instruments	107
6.6	Calibration of force standards and test instruments	120
	Bibliography	132
7	Measurement of temperature	133
7.1	Introduction	133
7.2	SI units	133
7.3	Thermodynamic scale	133
7.4	Practical temperature scales	134
7.5	International Temperature Scale of 1990 (ITS-90)	134
7.6	Industrial thermometers	135
7.7	Calibration of thermometers	154
	Bibliography	163
8	Electrical measurement standards	164
8.1	Introduction	164
8.2	SI units	164
8.3	Primary standards	166
8.4	Secondary standards	171
8.5	Working standards	180
8.6	Calibration of a multifunction calibrator	183
8.7	Calibration of multimeters and other instruments	184
	Bibliography	192
9	Uncertainty of measurements	193
9.1	Introduction	193

9.2	Basic concepts	193
9.3	Recommendations of the ISO guide	194
9.4	Examples of uncertainty calculations	198
	Bibliography	203
	Appendix 1	204
	Appendix 2	206
	Appendix 3	207
	Index	209

Requirements of ISO 9000 standards for test and measuring equipment

1.1 Introduction

Certification to ISO 9000 standards has become a primary requirement for both manufacturing and service-oriented organizations. Calibration and control of test, measurement and inspection equipment is one of the more important requirements given in the standard. A company that wants to obtain ISO 9000 certification therefore has to look into this vital aspect of their operations.

Test or calibration laboratories wishing to obtain independent third party certification should be guided by the requirements of ISO/IEC 17025: 1999 (formerly ISO/IEC Guide 25). A brief outline of the requirements of ISO 9001 standards is given in this chapter.

1.2 Evolution of ISO 9000 standards

The origin of the ISO 9000 series of quality management standards can be traced to the United States (US) military standards. The US military specifications MIL-I-Q9858 and MIL-I-45208 for quality inspection are the first standards to have specified requirements for quality assurance systems in the supplier's organization. Subsequently these standards were published as Allied Quality Assurance Publications (AQAP) 1, 4 and 9.

In 1972, the United Kingdom established UK Defence Standards 05/21, 05/24 and 05/29 based on the AQAP documents 1, 4 and 9. The famous British Standard BS 5750: 1979, parts 1, 2 and 3, were based on the presently obsolete UK Defence Standards 05/21, 05/24 and 05/29.

In 1985 the International Organization for Standardization through its Technical Committee on Quality Management and Assurance (ISO/TC 176) undertook the preparation of a series of international standards for quality management and BS 5750, which had been used successfully by the British Standards Institution for quality system certification, became the natural choice

for basing the new international standard. After much deliberation and arguments ISO 9001, ISO 9002 and ISO 9003 were published in 1987. These standards were then adopted by a significant number of national standards bodies, including the United Kingdom, and were published as their national standards. The ISO 9000 series was also published as a European Standard series EN 29000 by the European Committee on Standardization (CEN).

In 1994 a revision of the series was undertaken, and an updated and revised set of standards was published. In the mean time a large number of organizations obtained certification against ISO 9001 and ISO 9002 standards. The usefulness of the standards for quality assurance of products and services was beginning to be accepted worldwide, though there were some organizations that were not entirely convinced by the necessity of a documented quality system as required by the standards.

A further revision of the standards was undertaken during 1996 to 2000, and a revised and improved set of standards known as ISO 9000: 2000 has been published. In the new standard certification can be obtained only against the ISO 9001 standard. ISO 9002 and ISO 9003 standards have been withdrawn. ISO 9004 has been published as a complementary guidance document.

1.3 Requirements of ISO 9001: 2000

The requirements of the ISO 9001: 2000 standard in respect of test, inspection and measuring equipment are summarized below:

- (a) The organization shall identify the measurements to be made and the measuring and monitoring devices required to assure conformity of product to specified requirements.
- (b) Measuring and monitoring devices shall be used and controlled to ensure that measurement capability is consistent with the measurement requirements.
- (c) Measurement and monitoring shall be calibrated and adjusted periodically or prior to use, against devices traceable to international or national standards; where no such standards exist the basis used for calibration shall be recorded.
- (d) Where applicable measuring and monitoring devices shall:
 - (i) be safeguarded from adjustments that would invalidate the calibration;
 - (ii) be protected from damage and deterioration during handling, maintenance and storage;
 - (iii) have the results of their calibration recorded; and
 - (iv) have the validity of previous results reassessed if they are subsequently found to be out of calibration, and corrective action taken.

Some guidelines for achieving these requirements are given. The international standard ISO 10012 – Part 1 is also a useful source of information for quality assurance of measuring equipment.