



Clinical Applications of Digital Dental Technology



Edited by
Radi Masri
Carl F. Driscoll

WILEY Blackwell

Clinical Applications of Digital Dental Technology


Edited by

Radi Masri, DDS, MS, PhD

Associate Professor, Department of Endodontics, Prosthodontics, and Operative Dentistry,
School of Dentistry, University of Maryland, Baltimore, Maryland, USA

Carl F. Driscoll, DMD

Professor and Director, Prosthodontic Residency, Department of Endodontics, Prosthodontics, and Operative Surgery,
School of Dentistry, University of Maryland, Baltimore, Maryland, USA



WILEY Blackwell

This edition first published 2015 © 2015 by John Wiley & Sons, Inc.

Editorial offices:

1606 Golden Aspen Drive, Suites 103 and 104, Ames, Iowa 50010, USA
The Atrium, Southern Gate, Chichester, West Sussex, PO19 8SQ, UK
9600 Garsington Road, Oxford, OX4 2DQ, UK

For details of our global editorial offices, for customer services and for information about how to apply for permission to reuse the copyright material in this book please see our website at www.wiley.com/wiley-blackwell.

Authorization to photocopy items for internal or personal use, or the internal or personal use of specific clients, is granted by Blackwell Publishing, provided that the base fee is paid directly to the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923. For those organizations that have been granted a photocopy license by CCC, a separate system of payments has been arranged. The fee codes for users of the Transactional Reporting Service are ISBN-13: 978-1-1186-5579-5 / 2015.

Designations used by companies to distinguish their products are often claimed as trademarks. All brand names and product names used in this book are trade names, service marks, trademarks or registered trademarks of their respective owners. The publisher is not associated with any product or vendor mentioned in this book.

The contents of this work are intended to further general scientific research, understanding, and discussion only and are not intended and should not be relied upon as recommending or promoting a specific method, diagnosis, or treatment by health science practitioners for any particular patient. The publisher and the author make no representations or warranties with respect to the accuracy or completeness of the contents of this work and specifically disclaim all warranties, including without limitation any implied warranties of fitness for a particular purpose. In view of ongoing research, equipment modifications, changes in governmental regulations, and the constant flow of information relating to the use of medicines, equipment, and devices, the reader is urged to review and evaluate the information provided in the package insert or instructions for each medicine, equipment, or device for, among other things, any changes in the instructions or indication of usage and for added warnings and precautions. Readers should consult with a specialist where appropriate. The fact that an organization or Website is referred to in this work as a citation and/or a potential source of further information does not mean that the author or the publisher endorses the information the organization or Website may provide or recommendations it may make. Further, readers should be aware that Internet Websites listed in this work may have changed or disappeared between when this work was written and when it is read. No warranty may be created or extended by any promotional statements for this work. Neither the publisher nor the author shall be liable for any damages arising herefrom.

Library of Congress Cataloging-in-Publication Data

Clinical applications of digital dental technology / editors, Radi Masri, Carl F. Driscoll.

p. ; cm.

Includes bibliographical references and index.

ISBN 978-1-118-65579-5 (pbk.)

I. Masri, Radi, 1975- , editor. II. Driscoll, Carl F., editor.

[DNLM: 1. Radiography, Dental, Digital-methods. 2. Computer-Aided Design. 3. Radiation Dosage. 4. Technology, Dental. WN 230]

RK309

617.6'07572-dc23

2015007790

A catalogue record for this book is available from the British Library.

Wiley also publishes its books in a variety of electronic formats. Some content that appears in print may not be available in electronic books.

Cover image: [Production Editor to insert]

Cover design by [Production Editor to insert]

Typeset in 9.5/11.5pt Palatino by Laserwords Private Limited, Chennai, India.

Printed and bound in Singapore by Markono Print Media Pte Ltd

To family, near, and far.

Contributors

Nadim Z. Baba, DMD, MSD

Professor, Department of Restorative Dentistry,
Loma Linda University School of Dentistry,
Loma Linda, CA, USA

Francesca Bonino, DDS

Postgraduate resident, Department of
Periodontology, Tufts University School of
Dental Medicine, Boston, MA, USA

Jacinto A. Cano Peyro, DDS

Instructor, Department of Prosthodontics &
Operative Dentistry, Tufts University School of
Dental Medicine, Boston, MA, USA

Carl F. Driscoll, DMD

Professor and Director, Advanced Education in
Prosthodontics, Department of Endodontics,
Prosthodontics, and Operative Dentistry, School
of Dentistry, Maryland, Baltimore, MD, USA

Dennis J. Fasbinder, DDS

Clinical Professor, Department of Cariology,
Endodontics, and Restorative Services,
University of Michigan School of Dentistry,
Ann Arbor, MI, USA

Ashraf F. Fouad, BDS, DDS, MS

Professor and Chair, Department of Endodontics,
Prosthodontics and Operative Dentistry, School
of Dentistry, University of Maryland, Baltimore,
MD, USA

Charles J. Goodacre, DDS, MSD

Professor, Restorative Dentistry, Loma Linda
University, School of Dentistry, Loma Linda,
CA, USA

Gerald T. Grant DMD, MS

Captain, Dental Corps, United States Navy,
Service Chief, 3D Medical Applications Center,
Department of Radiology, Walter Reed
National Military Medical Center, Director of
Craniofacial Imaging Research, Naval
Postgraduate Dental School, Bethesda, MD,
USA

Gary D. Hack

Associate Professor and Director of Clinical
Stimulation, Department of Endodontics,
Prosthodontics, and Operative Dentistry,
School of Dentistry, University of Maryland,
Baltimore, MD, USA

Julie Holloway, DDS, MS

Professor and Head, Department of
Prosthodontics, The University of Iowa College
of Dentistry, Iowa City, IA, USA

Jason Jamali, DDS, MD

Clinical Assistant Professor, Department of Oral
and Maxillofacial Surgery, University of
Illinois, Chicago, IL, USA

Georgios Kanavakis, DDS, MS

Assistant Professor, Department of Orthodontics
and Dentofacial Orthopedics, Tufts University,

School of Dental Medicine, Boston,
MA, USA

Mathew T. Kattadiyil, BDS, MDS, MS, FACP

Director, Advanced Specialty Education Program
in Prosthodontics, Loma Linda University
School of Dentistry, Loma Linda, CA, USA

Joanna Kempler, DDS, MS

Clinical Assistant Professor, Department of
Endodontics, Prosthodontics and Operative
Dentistry, University of Maryland, Baltimore,
MD, USA

Antonia Kolokythas, DDS, MSc

Associate Professor, Program Director,
Department of Oral and Maxillofacial Surgery,
Multidisciplinary Head and Neck Cancer
Clinic, University of Illinois at Chicago,
Chicago, IL, USA

Radi Masri, DDS, MS, PhD

Associate Professor, Advanced Education in
Prosthodontics, Department of Endodontics,
Prosthodontics, and Operative Dentistry, School
of Dentistry, Maryland, Baltimore, MD, USA

Michael Miloro, DMD, MD, FACS

Professor and Head, Department of Oral and
Maxillofacial Surgery, University of Illinois at
Chicago, Chicago, IL, USA

Alexandra Patzelt, DMD, Dr med dent

Visiting Scholar, Department of Periodontics,
School of Dentistry, University of Maryland,
Baltimore, MD, USA

Sebastian B. M. Patzelt, DMD, Dr med dent

Associate Professor, Department of Prosthetic
Dentistry, Center for Dental Medicine, Medical
Center – University of Freiburg, Freiburg i. Br.,
Germany

Jeffery B. Price, DDS, MS

Associate Professor, Director of Oral &
Maxillofacial Radiology, Department of
Oncology & Diagnostic Sciences, University of
Maryland School of Dentistry, Baltimore, MD,
USA

Carroll Ann Troitman, BDS, MS, MA

Professor and Chair, Department of Orthodontics,
Tufts University School of Dental Medicine,
Boston, MA, USA

Hans-Peter Weber, DMD, Dr med dent

Professor and Chair, Department of
Prosthodontics and Operative Dentistry, Tufts
University School of Dental Medicine, Boston,
MA, USA

Foreword

Advances in technology have resulted in the development of diagnostic tools that allow clinicians to gain a better appreciation of patient anatomy that then leads to potential improvements in treatment options. Biomechanical engineering coupled with advanced computer science has provided dentistry with the ability to incorporate three-dimensional imaging into treatment planning and surgical and prosthodontic treatment. Optical scanning of tooth preparations and dental implant positions demonstrates accuracy that is similar to or possibly an improvement upon that seen with traditional methods used to make impressions and create casts.

For example, with this technology, orthodontic treatment can be reevaluated to assess outcomes. Today, orthodontic treatment can be planned and executed differently. With CT scanning on the orthodontic patient, dentists can better understand the boney limitation of a proposed treatment and timing of the treatment and dental implants can be used to create anchorage to move the teeth more easily. Every aspect of dentistry has been affected by digital technology, and in most instances, this has resulted in improvements of clinical treatment.

Restorative Dentistry and Prosthodontics are likely to experience the most dramatic changes relative to the incorporation of digital technology. Three-dimensional imaging provides the clinician with an ability to analyze bone quantity and quality that should lead to more effective development

of surgical guides. Likewise, hard and soft tissue grafting may be anticipated in advance, which will allow improved site development for esthetics and function. Such planning allows more affective provisionalization of the teeth and implants. By digitally understanding the design and tooth position, a provisional prosthesis can be fabricated using a monolithic premade block of acrylic, composite, or hybrid resin, thereby improving the ultimate strength of these prostheses. Dental material science has responded by producing materials that are more esthetic and can best provide a better potential for long-term survival and stability. Dental ceramics now can be milled on machines that can accept ever-improving algorithms to provide the most accurate prosthesis. Today, materials such as lithium disilicate, zirconia, and titanium are easily milled in machines that are self-calibrating and can eliminate the cuttings, so that accuracy is insured. In-office or in-laboratory CAD/CAM equipment is constantly improving, and it is clear that in years to come surgical guides and most types of ceramic restorations will be able to be produced accurately and predictably in the office environment. This will change some of the duties of the dental technologist but in no way will compromise the necessity of having these trained and very talented professionals more involved in designing, individualizing color and characterization, correcting marginal discrepancies, and refining the prosthetic occlusions that

are required. The dental technologist represents the most important function in delivering a restoration, that of quality control.

The future is exceedingly bright for all involved in the provision of dental care; moreover, the incorporation of digital dental procedures promises to improve care for the most important person in the treatment team, the patient.

The authors should be commended for bringing such valuable information and insight to the profession. At this point, information is what everyone most desire and one can be very proud of all the efforts forward-thinking professionals, engineers, and material scientists are bringing to the table. An honest appraisal of where we are today and what the potential future can be will drive the industry to create better restorative materials and engineered equipment and algorithms to dentistry.

Kenneth Malament

Preface

The evolution of the art and science of dentistry has always been gradual and steady, driven primarily by innovations and new treatment protocols that challenged the conventional wisdom such as the invention of the turbine handpiece and the introduction of dental endosseous implants.

While these innovations were few and far between, the recent explosion in digital technology, software, scanning, and manufacturing capabilities caused an unparalleled revolution leading to a major paradigm shift in all aspects of dentistry. Not only is digital radiography routine practice in dental clinics these days, but virtual planning and computer-aided design and manufacturing are also becoming mainstream. Digital impressions, digitally fabricated dentures, and the virtual patient are no longer science fiction, but are, indeed, a reality.

A new discipline, digital dentistry, has emerged, and the dental field is scrambling to fully integrate it into clinical practice and educational curriculums and as such, a comprehensive textbook that details the digital technology available and describes its indications, contraindications, advantages, disadvantages, limitations, and applications in the various dental fields is sorely needed.

There are a limited number of books and book chapters that address digital radiography, digital surgical treatment planning, and digital photography, but none address *digital dentistry* comprehensively. Although these topics will be addressed in this book, the scope is entirely different. The main focus is the practical application of digital technology in all aspects of dentistry. Available technologies will be discussed and critically evaluated to detail how they are incorporated in daily practice across all specialties. Realizing that technology changes rapidly, developing technologies and those expected to be on the market in the future will also be discussed.

Thus, this book is intended for a broad audience that includes dental students, general practitioners, and specialists of all the dental disciplines including prosthodontists, endodontists, orthodontists, oral and maxillofacial surgeons, periodontists, and oral and maxillofacial radiologists. It is also useful for laboratory technicians, dental assistants and dental hygienists, and anyone interested in recent digital advances in the dental field. We hope that the reader will gain a comprehensive understanding of digital applications in dentistry.

Contents

Contributors	ix		
Foreword	xi		
Preface	xiii		
1 Digital Imaging	1	8 Digital Design and Manufacture of Implant Abutments	167
Jeffery B. Price and Marcel E. Noujeim		Radi Masri, Joanna Kempler, and Carl F. Driscoll	
2 Digital Impressions	27	9 Digital Applications in Endodontics	177
Gary D. Hack, Ira T. Bloom, and Sebastian B. M. Patzelt		Ashraf F. Fouad	
3 Direct Digital Manufacturing	41	10 From Traditional to Contemporary: Imaging Techniques for Orthodontic Diagnosis, Treatment Planning, and Outcome Assessment	193
Gerald T. Grant		Georgios Kanavakis and Carroll Ann Trotman	
4 Digital Application in Operative Dentistry	57	11 Clinical Applications of Digital Dental Technology in Oral and Maxillofacial Surgery	207
Dennis J. Fasbinder and Gisele F. Neiva		Jason Jamali, Antonia Kolokythas, and Michael Miloro	
5 Digital Fixed Prosthodontics	75	12 The Virtual Patient	231
Julie Holloway		Alexandra Patzelt and Sebastian B. M. Patzelt	
6 CAD/CAM Removable Prosthodontics	107		
Nadim Z. Baba, Charles J. Goodacre, and Mathew Kattadiyil		Index	241
7 Digital Implant Surgery	139		
Hans-Peter Weber, Jacinto Cano, and Francesca Bonino			

1

Digital Imaging

Jeffery B. Price and Marcel E. Noujeim

Introduction

Imaging, in one form or another, has been available to dentistry since the first intraoral radiographic images were exposed by the German dentist, Otto Walkhoff (Langland *et al.*, 1984), in early 1896, just 14 days after W.C. Roentgen publicly announced his discovery of X-rays (McCoy, 1919; Bushong, 2008). Many landmark improvements have been made over the more than 115-year history of oral radiography.

The first receptors were glass, however, film set the standard for the greater part of the twentieth century until the 1990s, when the development of digital radiography for dental use was commercialized by the Trophy company who released the RVGui system (Mouyen *et al.*, 1989). Other companies such as Kodak, Gendex, Schick, Planmeca, Sirona, and Dexis were also early pioneers of digital radiography.

The adoption of digital radiography by the dental profession has been slow but steady and seems to have been governed, at least partly, by the “diffusion of innovation” theory espoused by Dr. Everett Rogers (Rogers, 2003). His work describes how various technological improvements have been adopted by the end

users of technology throughout the second half of the twentieth century and the early twenty-first century. Two of the most important tenets of adoption of technology are the concepts of threshold and critical mass.

Threshold is a trait of a group and refers to the number of individuals in a group who must be using a technology or engaging in an activity before an interested individual will adopt the technology or engage in the activity. Critical mass is another characteristic of a group and occurs at the point in time when enough individuals in the group have adopted an innovation to allow for self-sustaining future growth of adoption of the innovation. As more innovators adopt a technology such as digital radiography, the perceived benefit of the technology becomes greater and greater to ever-increasing numbers of other future adopters until eventually the technology becomes commonplace.

Digital radiography is the most common advanced dental technology that patients experience during diagnostic visits. According to one leading manufacturer in dental radiography, digital radiography is used by 60% of the dentists in the United States (Tokhi, J., 2013, personal communication). If you are still using film, the

question should not be “Should I switch to a digital radiography system?”, but instead “Which digital system will most easily integrate into my office?”

This leads to another question, what advantages does digital radiography offer the dental profession as compared to simply continuing with the use of conventional film? What are the reasons that increasing numbers of dentists are choosing digital radiographic systems over conventional film systems? Let us look at them.

Digital versus conventional film radiography

The most common speed class, or sensitivity, of intraoral film has been, and continues to be, D-speed film; the prime example of this film in the US market is Kodak’s Ultra-Speed (NCRP, 2012). The amount of radiation dose required to generate a diagnostic image using this film is approximately twice the amount required for Kodak’s Insight, an F-speed film. In other words, F-speed film is twice as fast as D-speed film. According to Moyal, who used a randomly selected survey of 340 dental facilities from 40 states found in the 1999 NEXT data, the skin entrance dose of a typical D-speed posterior bitewing is approximately 1.7 mGy (Moyal, 2007). Furthermore, according to the National Council on Radiation Protection and Measurements (NCRP) Report #172, the median skin entrance dose for a D-speed film is approximately 2.2 mGy while the typical E-F-speed film dose is approximately 1.3 mGy and the median skin entrance dose from digital systems is approximately 0.8 mGy (NCRP, 2012). According to NCRP Report #145 and others, it appears that dentists who are using F-speed film tend to overexpose the film and then under develop it; this explains why the radiation dose savings with F-speed film is not as great as it could be because F-speed film is twice as fast as D-speed film (NCRP, 2004; NCRP, 2012). If F-speed film were used per the manufacturers’ instructions, the exposure time and/or milliamperage (total mAs) would be half that of D-speed film and the radiation dose would then be half.

Why has there been so much resistance for dentists to move away from D-speed film and embrace digital radiography? First of all, operating a dental

office is much like running a fine-tuned production or manufacturing facility; dentists spend years perfecting all the systems needed in a dental office, including the radiography system. Changing the type of imaging system risks upsetting the dentist’s capability to generate comprehensive diagnoses; therefore, in order to persuade individual dentists to change, there has to be compelling reasons, and, until recently, most of the dentists in the United States have not been persuaded to make the change to digital radiography. It has taken many years to reach the threshold and the critical mass for the dental profession to make the switch to digital radiography. Moreover, in all likelihood, there are dentists today who will retire from active practice before they switch from film to digital.

There are many reasons to adopt digital radiography: decreased environmental burdens by eliminating developer and fixer chemicals along with silver and iodide bromide chemicals; improved accuracy in image processing; decreased time required to capture and view images, which increases the efficiency of patient treatment; reduced radiation dose to the patient; improved ability to involve the patient in the diagnosis and treatment planning process with co-diagnosis and patient education; and viewing software to dynamically enhance the image (Wenzel, 2006; Wenzel and Møystad, 2010; Farman *et al.*, 2008). However, if dentists are to enjoy these benefits, the radiographic diagnoses for digital systems must be at least as reliably accurate as those obtained with film (Wenzel, 2006).

Two primary cofactors seem to be more important than others in driving more dentists away from D-speed and toward digital radiography – the increased use of computers in the dental office and the reduced radiation doses seen in digital radiography. We will explore these factors further in the next section.

Increased use of computers in the dental office

This book’s focus is digital dentistry and later sections will deal with how computers interface with every facet of dentistry. The earliest uses of the computer in dentistry were in the business

office and accounting. Over the ensuing years, computer use spread to full-service practice management systems with digital electronic patient charts including digital image management systems. The use of computers in the business operations side of the dental practice allowed dentists to gain experience and confidence in how computers could increase efficiency and reliability in the financial side of their practices. The next step was to allow computers into the clinical arena and use them in patient care. As a component of creating the virtual dental patient, initially, the two most prominent roles were electronic patient records and digital radiography. In the following sections, we will explore the attributes of digital radiography including decreased radiation doses as compared to film; improved operator workflow and efficiency; fewer errors with fewer retakes; wider dynamic range; increased opportunity for co-diagnosis and patient education; improved image storage and retrievability; and communication with other providers (Farman *et al.*, 2008; Wenzel and Møystad, 2010).

Review of basic terminology

Throughout this section, we will be using several terms that may be new to you, especially if you have been using conventional film; therefore, we will include the following discussion of some basic oral radiology terms, both conventional and digital. Conventional intraoral film technology, such as periapical and bitewing imaging, uses a *direct* exposure technique whereby the X-ray photons directly stimulate the silver bromide crystals to create the latent image. Today's *direct digital* X-ray sensor refers most commonly to a complementary metal oxide semiconductor (CMOS) sensor that is directly connected to the computer via a USB port. At the time of the exposure, X-ray photons are detected by cesium iodide or perhaps gadolinium oxide scintillators within the sensor, which then emit light photons; these light photons are then detected within the sensor pixel by pixel, which allows for almost instantaneous image formation on the computer display. Most clinicians view this instantaneous image formation as the most advantageous characteristic of direct digital imaging.

The other choice for digital radiography today is an *indirect digital* technique known as photo-stimulable phosphor or PSP plates; these plates resemble conventional film in appearance and clinical handling. During exposure, the latent image is captured within energetic phosphor electrons; during processing, the energetic phosphors are stimulated by a red laser light beam; the latent energy stored in the phosphor electrons is released as a green light, which is captured, processed, and finally digitally manipulated by the computer's graphic card into images relayed to the computer's display. The "indirect" term refers to the extra processing step of the plates as compared to the direct method when using the CMOS sensor. The most attractive aspect of PSP may be that the clinical handling of the phosphor plates is exactly like handling film; so, most offices find that the transition to PSP to be very manageable and user-friendly.

Panoramic imaging commonly uses direct digital techniques as well. The panoramic X-ray beam is collimated to a slit; therefore, the direct digital sensor is several pixels wide and continually captures the signal of the remnant X-ray beam as the panoramic X-ray source/sensor assembly continually moves around the patient's head; the path of the source/sensor assembly is the same whether the receptor is an indirect film, PSP, or direct digital system. Clinicians who are using intraoral direct digital receptors generally opt for a direct digital panoramic system to avoid the need to purchase a PSP processor for their panoramic system.

Orthodontists require a cephalometric system and when moving from film to digital, again have two choices: direct digital and indirect digital. The larger flat panel digital receptor systems provide the instantaneous image but are slightly more costly than the indirect PSP systems; however, the direct digital systems obviate the need to purchase and maintain PSP processors. The higher the volume of patients in the office, the quicker is the financial payback for the direct digital X-ray machine.

Image quality comparison between direct and indirect digital radiography

Some dentists will make the decision of which system to purchase based solely on the speed of the system, with the direct digital system being the fastest. There are other factors as well: dentists often ask about image quality. Perhaps the better question to ask may be, "Is there a significant difference between the diagnostic capability of direct and indirect digital radiography systems?" One of the primary diagnostic tasks facing dentists on a daily basis is caries diagnosis, and there are several studies that have evaluated the efficacy of the two systems at this common task. The answer is that there is no difference between the two systems in diagnostic efficacy – either direct digital or indirect digital with PSP plates will diagnose caries equally well, in today's modern systems (Wenzel *et al.*, 2007; Berkhout *et al.*, 2007; Li *et al.*, 2007).

One important consideration to consider when comparing systems is to make sure that the images have the same *bit depth*. Bit depth refers to the numbers of shades of gray used to generate the image and are expressed exponentially in Table 1.1.

The early digital systems had a bit depth of 8 with 256 shades of gray, which may seem fine because the human eye can only detect approximately 20 to 30 shades of gray at any one time in any one image; however, most digital systems today generate images at 12 or even 16 bit depth, that is, images that have 4,096 to 65,536 shades of gray (Russ, 2007). Proper image processing is a skill that must be learned in order to fully utilize all of the information contained in today's digital images. Conventional film systems do not have discrete shades of gray; rather, film systems are analog and have an infinite number of possible shades of gray depending only on the numbers of silver atoms activated in each cluster of silver atoms in the latent image within the silver halide lattice of the film emulsion. Therefore, when comparing systems, ensure that the bit depth of the systems is comparable; and, remember that over time, the higher bit depth systems will require larger computer storage capacities due to the larger file sizes associated with the increased amount of digital information requirements of the larger bit depth images. It is expected that

Table 1.1 Bit depth table that gives the relation of the exponential increase in the number of shades of gray available in images as the bit depth increases.

Bit depth	Expression	Number of shades of gray
1	2^1	2
2	2^2	4
3	2^3	8
4	2^4	16
5	2^5	32
6	2^6	64
7	2^7	128
8	2^8	256
9	2^9	512
10	2^{10}	1024
11	2^{11}	2048
12	2^{12}	4096
13	2^{13}	8192
14	2^{14}	16384
15	2^{15}	32768
16	2^{16}	65536

in the future, most systems will use images of a minimum of 12 bit depth quality and many are already using images of 16 bit depth quality.

Amount of radiation required to use direct and indirect digital radiography

One other factor that dentists should consider when evaluating which system to use is how much radiation is required for each system to generate a diagnostic image. In order to determine the answer to this question, clinicians should be familiar with the term *dynamic range*, which refers to the performance of a radiographic receptor system in relation to the amount of radiation required to produce a desired amount of optical density within the image. The Hurter and Driffield (H&D) characteristic curve chart was initially developed for use with film systems and can also be used with direct digital and indirect digital systems

(Bushong, 2008; Bushberg *et al.*, 2012). The indirect digital system with PSP plates has the widest dynamic range, even wider than film, which means that PSP plates are more sensitive to lower levels of radiation than either conventional film or direct digital CMOS detectors; and, at the upper range of diagnostic exposures, the PSP plates do not experience burnout as quickly as film or direct digital until very high radiation doses are delivered. This means that the PSP system can handle a wider range of radiation dose and still deliver a diagnostic image, which may be a good feature, but for patient safety, this may be a negative feature because dentists may consistently be unaware that the operator of the equipment is delivering higher radiation doses than are necessary simply because their radiographic system has not been calibrated properly (Bushong, 2008; Bushberg *et al.*, 2012; Huda *et al.*, 1997; Hildebolt *et al.*, 2000).

Radiation safety of digital radiography

There are several principles of radiation safety: ALARA, justification, limitation, optimization, and the use of selection criteria. We will briefly review these and then discuss how digital radiography plays a vital role in the improved safety of modern radiography.

The acronym *ALARA* stands for As Low As Reasonably Achievable and, in reality, is very straightforward. In the dental profession, dental auxiliaries and dental professionals are required to use medically accepted radiation safety techniques that keep radiation doses low and that do not cause an undue burden on the operator or clinician. An example from the NCRP Report #145 Section 3.1.4.1.4 states “Image receptors of speeds slower than ANSI Speed Group E *shall not* be used for intraoral radiography. Faster receptors *should* be evaluated and adopted if found acceptable” (NCRP, 2004). This means that offices do not have to switch to digital but rather could switch to E- or F-speed film but *must* switch to at least E-speed film in order to be in compliance with this report. This is but one example of practicing ALARA. In the United States, federal and nationally recognized agencies such as the Food and Drug Administration (FDA)

and the NCRP issue guidelines and best practice recommendations; however, laws are enforced on the state level, which results in a confusing patchwork of various regulations, and dentists sometimes confuse what must be done with what should be done, especially because a colleague in a neighboring state must follow different laws. For example, although it is recommended by the NCRP but not legally required in many states, the state of Maryland now legally requires dentist to practice ALARA (Maryland, 2013), although the neighboring state of Virginia does not specifically require this in their radiation protection regulations (Commonwealth of Virginia, 2008). Therefore, in the state of Maryland, in order to satisfy legal requirements, dentists will soon be replacing D-speed film with either F-speed film or digital systems. Internationally, groups such as the International Commission on Radiological Protection (ICRP), the United Nations Scientific Committee on the Effects of Atomic Radiation (UNSCEAR), and the Safety and Efficacy in Dental Exposure to CT (SEDENTEXTCT) have provided well-researched recommendations on the use of imaging in dentistry and guidance on the information of the effects of ionizing radiation on the human body (ICRP, 1991; Valentin, 2007; Ludlow *et al.*, 2008; UNSCEAR, 2001; Horner, 2009).

When a clinician goes through the process of examining a patient and formulating a diagnostic question, he or she is justifying the radiographic examination. This principle of *justification* is one of the primary principles of radiation safety. With digital radiography, our radiation doses are very low: so low, in fact, that if we have a diagnostic question that can only be answered with the information obtained from a dental radiograph, the risk from the radiograph is low enough that the “risk to benefit analysis” is always in favor of exposing the radiograph. There will always be enough of a benefit to the patient to outweigh the very small risk of the radiographic examination, as long as there is significant diagnostic information to be gained from the X-rays.

The principle of *limitation* means that the X-ray machine operator is doing everything possible to limit the actual size of the X-ray beam: that is, collimation of the X-ray beam. For intraoral radiography, rectangular collimation is recommended for routine use by the NCRP and there are

various methods available to achieve collimation of the beam. Rectangular collimation reduces the radiation dose to the patient by approximately 60%. In panoramic imaging, the X-ray beam is collimated to a slit-shape. Moreover, in cone-beam CT, the X-ray beam has a cone shape.

In late 2012, the FDA and ADA issued the latest recommendations for selection criteria of the dental patient. These guidelines give the dentist several common scenarios that are seen in practice and offer suggestions on which radiographs may be appropriate. This article provides an excellent review of the topic and is best summarized by a sentence found in its conclusion: "Radiographs should be taken only when there is an expectation that the diagnostic yield will affect patient care" (ADA & FDA, 2012).

How does digital radiography assist with managing radiation safety? As mentioned earlier, digital receptors require less radiation dose than film receptors. In the 2012 NCRP Report #172, section 6.4.1.3, it is recommended that US dentists adopt a diagnostic reference level (DRL) for intraoral radiographs of 1.2 mGy. This dose is the median dose for E- and F-speed film systems, and it is higher than the dose for digital systems. This means that in order to predictably achieve this ambitious goal, US dentists who are still using D-speed film will need to either switch to F-speed film or transition to a digital system (NCRP, 2012).

Radiation dosimetry

The dental profession owns more X-ray machines than any other profession; and, we expose a lot of radiographs. Our doses are very small, but today our patients expect us to be able to educate them and answer their questions about the safety of the radiographs that we are recommending and it is part of our professional responsibility to our patients. Let's review some vocabulary first. The International System uses the *Gray* (Gy) or milliGray (mGy), and microGray (μ Gy) to describe the amount of radiation dose that is absorbed by the patient's skin (skin entrance dose) or by their internal organs. This dose is measured by devices such as ionization chambers or optically stimulated dosimeters (OSLs). There

are different types of tissues in our body and they all have a different response or sensitivity to radiation; for instance, the child's thyroid gland seems to be the most sensitive tissue that is in the path of our X-ray beams while the mature mandibular nerve may be the least sensitive tissue type in the maxillofacial region (Hall and Giaccia, 2012). Of course, we only deal with diagnostic radiation, but there are other types of radiation such as gamma rays, alpha particles, and beta particles; in order to provide a way to measure the effect on the body's various tissues when exposed by radiation from the various sources, a term known as *equivalent dose* is used. This term is expressed in *Sieverts* (S) or milliSieverts (mSv), and microSieverts (μ Sv). Finally, another term known as *effective dose* is used to compare the risk of radiographic examinations. This is the most important term for dental professionals to be familiar with as this is the term that accounts for the type of radiation used (diagnostic in our case) and the type of tissues exposed by the X-ray beam in the examination, whether it is a bitewing, a panoramic, a cone beam CT or a chest X-ray, and so on. Using this term is like comparing apples with apples. By using this term, we can compare the risk of a panoramic radiograph with the risk of an abdominal CT or a head CT and so on.

When patients ask us about how safe a particular radiographic examination may be, they are really asking whether that X-ray is going to cause a fatal cancer. Moreover, when medical physicists estimate the risk of X-rays in describing effective dose as measured in Sieverts and microSieverts for dentistry, they are talking about the risk of developing a fatal cancer. The risk is usually given as the rate of excess cancers per million. In order to accurately judge this number, the clinician needs to know the background rate of cancer (and fatal cancer) in the population. According to the American Cancer Society, the average person, male or female, in the population of the United States has a 40% chance of developing cancer during his or her lifetime; furthermore, the rate of fatality of that group is 50%; therefore, the overall fatal cancer rate in the United States is 20%, or 200,000 per million people (Siegel *et al.*, 2014). Now, when you read in the radiation dosimetry

table (Table 1.2) that if a million people had a panoramic exposure and the excess cancer rate in those one million people was 0.9 per million, you will know that the total cancer rate changed from 200,000 per million to 200,000.9 per million. On a percentage basis, that is very small indeed – a 0.00045% risk of developing cancer. Of course, these are population-based numbers and are the best estimates groups like the NCRP can come up with, and you should also know that a very generous safety factor is built in. At the very low doses of ionizing radiation seen in most dental radiographic examinations experts such as medical physicists and molecular biologists do not know the exact mechanisms of how the human cell responds to radiation. So, to be safe and err on the side of caution, which is the prudent course of action, we all assume that some cellular and some genetic damage is possible due to a dose–response model known as the *linear no-threshold* model of radiation interaction, which is based on the assumption that in the low dose range of radiation exposures, any radiation dose will increase the risk of excess cancer and/or heritable disease in a simple proportionate manner (Hall and Giaccia, 2012).

There is one more column in Table 1.2 that needs some explanation – background equivalency. We live in a veritable sea of ionizing radiation, and the average person in the United States receives approximately $8\mu\text{Sv}$ of effective dose of ionizing radiation per day (NCRP, 2009). Take a look at the first examination – panoramic exposure; it has an effective dose of approximately $16\mu\text{Sv}$; if you divide $16\mu\text{Sv}$ by $8\mu\text{Sv}$ per day, the result is 2 days of background equivalency. Using this method, you now know that the amount of effective dose in the average panoramic examination equals the same amount of radiation that the average person receives over the course of 2 days. This same exercise has been completed for the examinations listed in the table; and, for examinations not listed, you can calculate the background equivalency by following the aforementioned simple calculations. The intended use of effective dose is to compare population risks; however, this use as described earlier is a quick and easy patient education tool that most of our patients can quickly understand.

Uses of 2D systems in daily practice

The use of standard intraoral and extraoral imaging for clinical dentistry have been available for many years and include caries and periodontal diagnosis, endodontic diagnosis, detection, and evaluation of oral and maxillofacial pathology and evaluation of craniofacial developmental disorders.

Caries diagnosis

The truth is that diagnosing early carious lesions with bitewing radiographs is much more difficult than it appears to be than at first impression. Most researchers have found that a predictably accurate caries diagnosis rate of 60% would be very acceptable in most studies. In a 2002 study, Mileman and van den Hout compared the ability of Dutch dental students and practicing general dentists to diagnose dentinal caries on radiographs. The students performed almost as well as the experienced dentists (Mileman and Van Den Hout, 2002; Bader *et al.*, 2001; Bader *et al.*, 2002; Dove, 2001). We will explore caries diagnosis and how modern methods of caries diagnosis are changing the paradigm from the past ways of diagnosing caries (Price, 2013).

Caries detection is a basic task that all dentists are taught in dental school. In principle, it is very simple – detect mineral loss in teeth visually, radiographically, or by some other adjunctive method. There can be many issues that affect this task, including training, experience, and subjectivity of the observer; operating conditions and reliability of the diagnostic equipment; these factors and others can all act in concert and often, the end result is that this “simple” task becomes complex. It is important to realize that the diagnosis of a carious lesion is only one aspect of the entire management phase for dental caries. In fact, there are many aspects of managing the caries process besides diagnosis. The lesion needs to be assessed as to whether the caries is limited to enamel or if it has progressed to dentin. A determination of whether the lesion progressed to a cavity needs to be made because a cavitated lesion will continue to trap plaque and will need to be restored. The activity level of the lesion